

TEST FACILITY

NAMSA
6750 Wales Road
Northwood, OH 43619
419.666.9455

CLIENT

SMOOTH-ON, INC
2000 Saint John Street
Easton, PA 18042

| | |
|-------------------------------------|--------------------------------|
| Test Report No: 14T_56435_03 | Date: December 29, 2014 |
|-------------------------------------|--------------------------------|

SAMPLE ID: The client identified the following test material as “**Dragon Skin**”

SAMPLING DETAIL: Test samples were submitted to the laboratory directly by the client. No special sampling conditions or sample preparation were observed by NAMSA.

DATE OF RECEIPT: Samples were received at NAMSA facilities on October 21, 2014.

TESTING PERIOD: November 10, 2014

AUTHORIZATION: Signed work order 116924.0 signed by Rodney C. Conn

TEST REQUESTED: Evaluate for the potential to elicit delayed dermal contact sensitization. This study was conducted based on the requirements of ISO 10993-10, Biological evaluation of medical devices, Part 10: Tests for irritation and skin sensitization.

TEST RESULTS: The test article showed no evidence of causing delayed dermal contact sensitization.

Prepared For:

Rodney C. Conn
Technical Compliance Specialist
SMOOTH-ON, INC.